

01679 2020 AnnRpt-LF



**REGIONAL DISTRICT
of Fraser-Fort George**

**2020 ANNUAL GREENHOUSE GAS EMISSION REPORT
FOOTHILLS BOULEVARD REGIONAL LANDFILL
AND COMPOST FACILITY
PRINCE GEORGE, BRITISH COLUMBIA
(JANUARY 1, 2020 – DECEMBER 31, 2020)**

March 2020

Prepared by:

**THE REGIONAL DISTRICT OF FRASER-FORT GEORGE
155 George Street
Prince George, British Columbia
V2L 1P8**

Darwin Paton
Environmental Technician

Petra Wildauer
General Manager Environmental Services

TABLE OF CONTENTS

1.0 INTRODUCTION 1

2.0 OBJECTIVE..... 2

3.0 SITE OVERVIEW..... 3

 3.1 Background.....3

 3.2 Site Description.....3

4.0 PROJECT OVERVIEW..... 4

 4.1 Well Field Overview4

 4.2 Control Plant Overview.....5

 4.3 Field Instrumentation and Datalogger.....5

 4.3.1 Datalogger.....6

 4.3.2 Gas Analyzer.....6

 4.3.3 Flow Meter.....6

5.0 LANDFILL GAS DATA SUMMARY 7

 5.1 Landfill Gas Collection.....7

 5.2 Landfill Gas Destruction7

 5.3 Landfill Management Facility Collection Efficiency7

6.0 MAINTENANCE AND SHUTDOWNS 8

 6.1 Scheduled Maintenance.....8

 6.2 Unscheduled Shutdowns.....8

 6.3 System Improvements.....9

 6.4 Proposed Facility Modifications and Alterations9

FIGURES

- Figure 3.1 Site Location Map
- Figure 3.2 Site Plan
- Figure 4.1 Existing LFG Management Facility

TABLES

- Table 5.1 Landfill Gas Collection Summary (2020)
- Table 5.2 Landfill Gas Destruction Summary (2020)
- Table 6.1 Summary of Scheduled Maintenance Shutdowns (2020)
- Table 6.2 Summary of Unscheduled Shutdowns (2020)

APPENDICES

- Appendix A Methane and Oxygen Analyzer Calibration Certificates
- Appendix B Flow Meter Calibration Certificate
- Appendix C Calculation Summary
- Appendix D Data Records

1.0 INTRODUCTION

The Regional District of Fraser-Fort George (RDFFG) is pleased to submit the following report entitled “2020 Annual Greenhouse Gas Emission Report” for the greenhouse gas emission reductions associated with the collection and destruction of landfill gas at the Foothills Boulevard Regional Landfill and Compost Facility.

The 2020 Annual Greenhouse Gas Emission Report presents the emission reductions associated with the capture and destruction of landfill gas at the Site from January 1, 2020 to December 31, 2020.

2.0 OBJECTIVE

The objective of this report is to present a summary of total landfill gas collected at the Site during the report period in compliance with the British Columbia Landfill Gas Management Regulation. This report presents the following information as prescribed in Section 14 of the aforementioned regulation:

- Quantity and composition of landfill gas collected at the site;
- Quantity and composition of landfill gas flared or used as an alternative to flaring;
- Maintenance and shutdown of landfill gas management facility;
- Landfill gas management facility efficiency; and
- Proposed modifications and alterations to the landfill gas management facility.

3.0 SITE OVERVIEW

3.1 Background

The Site was operated by the City of Prince George from 1976 until 1994. At that time, the RDFFG assumed responsibility for the Site, as recommended in the Regional Solid Waste Management Plan. Since that time, the RDFFG has made several improvements to the operation of the Site, including the addition of an on-site transfer station, a yard waste composting facility, a scale house, and a landfill gas collection system. In 2008 the Regional District of Fraser-Fort George obtained title to the landfill property from the province.

The Site serves a population of approximately 96,000 people and receives approximately 96 percent of the regional solid waste stream. The Site accepts waste from the Prince George municipal and commercial collection services, the general public, local institutions and businesses, and numerous regional transfer stations.

The property boundary for the Site encompasses an area of approximately 86.7 hectares. The landfill, composting, and recycling activities conducted at the Site encompass an area of approximately 25 hectares within the permitted landfill property.

3.2 Site Description

The Site is located in the northwest portion of the City of Prince George, British Columbia at 6595 Foothills Boulevard, northwest of the intersection with West Austin Road. The legal description of the Site is Block A of the northwest $\frac{1}{4}$ of District Lot 4053 and Block A of the northwest $\frac{1}{4}$ of District Lot 4048, Cariboo District. The location of the Site is shown on Figure 3.1.

The Site property (Figure 3.2) is bounded to the east by Foothills Boulevard and to the south by a gravel pit maintained and operated by the BC Ministry of Transportation and Infrastructure. The Site is bounded to the north and west by naturalized northern coniferous forest. A residential area, referred to as the Hart Highlands area of Prince George, is located approximately 500 metres east of the site, and a sport and recreation facility is located approximately 250 metres east of the site entrance. The Site is located approximately 2.5 kilometres north of the Nechako River.

4.0 PROJECT OVERVIEW

The following section presents an overview of the existing landfill gas management facility. The facility is comprised of a well field consisting of vertical extraction wells, horizontal collection pipes and control plant as illustrated in Figure 4.1. The original landfill gas collection system was installed and commissioned in 2002 as part of the closure of the eastern portion of Phase 1. During the summer of 2011 an additional eight (8) vertical wells were installed and commissioned to the west of the original sixteen (16) wells. They were completed with the tie-in of a 150 millimetre (mm) sub-header into the main landfill gas header as illustrated in Figure 4.1. Further wellfield expansion will be completed in 2021 to offset the decommissioning of the original 16 wells to implement a landfill remediation project.

4.1 Well Field Overview

The original 2002 landfill gas well field was composed of sixteen (16) vertical landfill gas extraction wells, connected to a main header via five (5) lateral lines. The original vertical extraction wells were constructed using CES-Landtec Accu-Flo wellheads, non-telescoping perforated PVC well assemblies, 100 mm diameter high density polyethylene (HDPE) laterals, and a 200 mm diameter HDPE main header.

Eight (8) new temporary extraction wells were constructed during the summer of 2011 using telescopic perforated PVC well assemblies, 50 mm in diameter. The landfill gas is transported through 100 mm diameter HDPE laterals, and a 150 mm diameter HDPE sub-header to the main landfill gas header. Butterfly isolation valves with buna seals, (150 mm and 200 mm respectively) were installed on both the main header and the new sub-header landfill gas lines.

Two (2) separate 150 mm diameter HDPE perforated telescopic horizontals (approximately 150 m in length and 60 m in length) were installed for the future collection of landfill gas.

The original main header pipe conveys landfill gas to the Control Plant over a distance of approximately 926 m. Condensate is managed via two (2) condensate knock-outs located adjacent to the Control Plant (condensate knockout #2) and adjacent to the public transfer station area (condensate knockout #1). The knockouts include electrically powered pumps. Condensate is conveyed from the knockouts via a force main to the leachate recirculation sump, located adjacent to condensate knockout #1. The condensate force main is located in the main header trench, adjacent to the 200 mm header pipe.

In the summer of 2013, an additional 580 metres of 200 mm diameter HDPE main header pipe was constructed off the west end of the existing main header. A 200 mm ball valve for isolation was installed where the main header was joined. Three (3) – 150 mm diameter HDPE sub-headers were then connected to the main header using 200 mm to 150 mm reducing tees and three (3) – 150 mm ball valves. These sub-headers were then flanged to two (2) – 150 mm horizontal collector pipes that were installed in the fall of 2011. A new in-line condensate trap with a pneumatically driven SP4000 condensate pump was also installed in the main header at this time.

In October of 2014, an additional 150 mm diameter HDPE perforated telescopic horizontal (approximately 250 m in length) was installed for the future collection of landfill gas.

In the autumn of 2016, a 350 m landfill gas header extension and perimeter road construction project was initiated in order to connect to the perforated horizontal collector pipe that was installed in 2014. One additional pneumatically driven SP4000 condensate pump was installed at this time. This project was completed in the early summer of 2017. During the summer of 2018 the last connection of the sub-header to the LFG Main Header pipe was completed.

2020 saw three (3) major projects undertaken that will increase the overall collection efficiency of the landfill gas system once completed.

In May of 2020, the final LFG Header Extension Project was completed. This extension completed a perimeter loop of the 200mm HDPE LFG Main Header pipe that was started with the initial LFG Project installed in 2002. The final length of the 200 mm HDPE LFG Main Header pipe is just over 2,000 meters. Its completion will greatly reduce variations of vacuum pressure throughout the collection system.

In the Autumn of 2020, the Remediation of the Phase 1, Stage one commenced with the decommissioning of five (5) of the original 2002 Landfill gas collection wells and two (2) of the Phase 1 leachate recirculation fields. (Collected leachate is, as of May 2020, conveyed via a 150mm force main to the municipal waste water collection system).

This remediation project was undertaken after it was identified that the air space available for solid waste disposal was significant to the landfills projected site life and the long-term increase in the collection of LFG. The remediation project will continue through 2021 and 2022 with the decommissioning of four (4) additional 2002 Landfill gas collection wells in the late spring of 2020.

Concurrently with the aforementioned remediation project a LFG Wellfield Expansion was undertaken in the south west area of the landfill. This well field expansion will see the installation of sixteen (16) new extraction wells using telescopic perforated PVC well assemblies, 50 mm in diameter with 100mm laterals and 150mm sub-headers connected to the now completed 200mm HDPE LFG Main Header perimeter loop. This project will be completed and the new wells brought on line in the spring of 2021.

4.2 Control Plant Overview

The landfill gas Control Plant is located at the northern extent of the landfill property. The landfill gas Control Plant is comprised of two primary components: abstraction building, and enclosed flare. The abstraction building includes three (3) main areas: blower room, electrical room, and garage facility. The blower room includes parallel National Turbine multi-stage blowers (2), isolation valve, condensate knock-out pot, landfill gas analyzer, and landfill gas flow meter. The electrical room houses the flare controls, programmable logic controller, blower variable frequency drive controls, and datalogger.

The flaring mechanism is an enclosed flare unit with a reported thermal capacity of 25 million BTU per hour. Hence, at 50 percent methane (volumetric basis), the flare has a maximum flow capacity of approximately 1,360 cubic metres per hour (800 cfm).

4.3 Field Instrumentation and Datalogger

Field instrumentation installed in the Control Plant associated with the monitoring of landfill gas include the following:

- Datalogger;
- Landfill Gas Analyzer; and
- Flow meter.

4.3.1 Datalogger

Data collected by the Control Plant programmable logic controller (PLC) is stored using Wonderware Historian and Information Server Portal at a rate of one (1) data sample every five (5) seconds. The following fourteen (14) channels are recorded by the PLC/Wonderware Historian software:

1. Air Damper Position Feedback;
2. Flare lower Thermocouple Temperature Feedback;
3. Flare middle Thermocouple Temperature Feedback;
4. Flare upper Thermocouple Temperature Feedback;
5. Blower #1 Speed Setpoint from HMI;
6. Blower #2 Speed Setpoint from HMI;
7. Blower #1 Speed Readback;
8. Blower #2 Speed Readback;
9. Discharge Gas Differential Pressure Readback;
10. Gas Discharge Flow Readback;
11. Gas Discharge Temperature Readback;
12. Gas Analyzer Methane Level Readback;
13. Gas Analyzer Oxygen Level Readback; and
14. Gas Discharge Pressure Readback.

4.3.2 Gas Analyzer

The volumetric methane concentration of the landfill gas is measured at the landfill gas Control Plant via a Galvanic Applied Sciences Inc. gas analyzer. The gas analyzer measures the volumetric concentration of methane using infrared absorption, and oxygen using a paramagnetic sensor. The reported accuracy range of the methane sensor is plus or minus 2 percent. A copy of the manufacturer's attestation of the accuracy range of the methane analyzer and gas analyzer calibration documentation for the reporting period is presented in Appendix A.

4.3.3 Flow Meter

The landfill gas flow rate is measured at the landfill gas Control Plant by a thermal flow meter (Thermal Instrument Company model 62-9) located in the landfill gas exhaust pipe, downstream of the blowers. The reported accuracy range of the flow meter is plus or minus 1 percent. A copy of the calibration records and manufacturer's attestation of the accuracy range of the flow meter is presented in Appendix B.

5.0 LANDFILL GAS DATA SUMMARY

The following section presents a summary of landfill gas collected and destroyed by the Site's landfill gas management facility. A summary of formulas and parameters utilized herein to calculate the mass of methane destroyed is presented in Appendix C.

5.1 Landfill Gas Collection

A summary of landfill gas collected at the site is summarized in Table 5.1. The raw data from which this summary is derived is presented in Appendix D. The total volume of landfill gas collected during the report period is 3,293,511 cubic metres. The average methane content of the collected landfill gas during the report is 55 percent (volumetric basis).

The landfill gas management facility operated for 7,416 hours 4 minutes during the report period with an 84 percent operational time.

5.2 Landfill Gas Destruction

The calculations presented herein are based upon the following parameters:

- Methane density of 0.667 kg per cubic metre at 25 °C;
- Flare hydrocarbon destruction efficiency of 98 percent; and
- Methane global warming potential of 25 times that of carbon dioxide.

A summary of landfill gas destroyed during the reporting period is presented in Table 5.2. A total of 3,227,641 cubic metres of landfill gas were flared during the report period, resulting in the destruction of 1,148 tonnes of methane. The equivalent mass of carbon dioxide destroyed during the reporting period is 28,703 tonnes.

5.3 Landfill Management Facility Collection Efficiency

The landfill gas management facility collection efficiency is estimated based upon the predicted average landfill gas generation rate for the Site (during the reporting period). Based upon the above stated total volume of landfill gas collected during the reporting period (3,293,411 cubic metres) and an estimated total landfill gas production volume of 13,515,853 cubic metres for the reporting period (Landfill Gas Generation Assessment Report, June, 2018) the operational collection efficiency of the landfill management facility is estimated to be 24 percent.

6.0 MAINTENANCE AND SHUTDOWNS

The following section presents a summary of shutdowns which occurred during the reporting period. In addition, a summary of scheduled modifications and expansion of the landfill gas management facility is presented herein.

6.1 Scheduled Maintenance

Scheduled maintenance shutdowns of the landfill gas management facility are summarized in Table 6.1. The facility was shut down for a total of 705 hours 18 minutes during the reporting period for scheduled maintenance.

Routine maintenance was carried out as scheduled in 2020. The semi-annual and annual maintenance were completed in October. Instrument calibrations were performed as scheduled during the 2020 reporting period.

The annual compressor inspection and maintenance was carried out in February.

In May the plant was shut down to facilitate the completion of the 2020 LFG Header Extension.

As part of the continuous maintenance program Blowers 1 and 2 underwent bearing and seal replacement and adjustment in July.

In September, the plant was shut down to accommodate the undertaking of gas well decommissioning in support of the Cell 1 remediation project.

All four Condensate Traps were serviced in October and were found to be in good operational condition.

In October, the Annual Environmental Sensor calibrations were carried out and completed.

The Landfill Gas System Annual Fire Alarm testing and maintenance was carried out in December.

Construction project considerations represent the main reason of decreased system efficiency in relation to scheduled shutdowns over the 2020 landfill gas collection period.

6.2 Unscheduled Shutdowns

Unscheduled shutdowns of the landfill gas management facility are summarized in Table 6.2. The facility was shut down for a total of 663 hours 9 minutes during the reporting period for unscheduled maintenance, repairs, and shutdowns due to control plant critical alarm conditions.

In February, March, July, August, October and November, local power supply failures caused the facility to shut down.

Further interruptions were encountered throughout the year when wildlife interrupted the power supply.

In August and October, instrument failure, due to a suspected power surge at the Abstraction Plant, shut down the system requiring repair.

In November, an Oxygen Analyzer failure shut down the system requiring examination and repair.

Power Supply interruptions represent the main reason of decreased system efficiency in relation to unscheduled shutdowns over the 2020 landfill gas collection period.

6.3 System Improvements

In May, the final LFG Header Extension Project was completed. This extension completed a perimeter loop of the 200 mm HDPE LFG Main Header pipe that was started with the initial LFG Project installed in 2002.

In the Autumn of 2020 a LFG Well Field Expansion Project was started that will see 16 additional LFG wells online by spring or early summer 2021.

6.4 Proposed Facility Modifications and Alterations

In 2021, in addition to the scheduled maintenance, the Regional District will be looking at replacing or upgrading key system components due to aging infrastructure as part of a continuous program to improve system reliability.


In early 2021 a Hydro realignment project will be undertaken to minimize power lines on site along with the installation of safe roosts and wildlife deterrents on required lines to help reduce power interruption.

Process Instrumentation upgrades will be undertaken and installed in the summer/fall of 2021 to help facilitate any potential alternative end use activities.

The Landfill Gas Well Field Expansion that was started in the fall of 2020 will be completed in the spring of 2021. Further progress on the remediation project that, once complete, will increase the overall efficiency of the current landfill gas collection system.

FIGURES

Figure 3.1



0 50 100m 200m
 SCALE 1:1250
 ALL DISTANCES IN METRES
 CONTOUR INTERVAL = 1.0m



REFERENCE
 Source: DMTI Spatial
 Datum: NAD 83 Projection: UTM Zone 10



 REGIONAL DISTRICT of Fraser-Fort George ENVIRONMENTAL SERVICES	155 GEORGE STREET PRINCE GEORGE, BC V2L 1P8 PHONE: (250) 960-4400 FAX: (250) 562-8626 TOLL FREE: 1-800-667-1959	DRAWN BY: E. MUDLE DATE: YY/MM/DD SCALE: 1:1 DWG. NO.: FIGURE 3.1
	PROJECT	
	FOOTHILLS BOULEVARD REGIONAL LANDFILL	
	DWG. TITLE SITE LOCATION PLAN	

Figure 3.2



Figure 4.1



TABLES

Table 5.1 Landfill Gas Collection Summary (2020)

Month	Volume LFG (m³)	Average Methane Content (%)	Operational Hours (hours)	Percent Time Operational (%)
January	323,565	53	744 hrs 0 min	100
February	237,616	55	542 hrs 31 min	78
March	314,257	54	740 hrs 0 min	99
April	221,845	55	510 hrs 23 min	71
May	177,800	55	426 hrs 32 min	57
June	337,551	55	704 hrs 18 min	98
July	334,405	54	691 hrs 14 min	93
August	351,000	51	696 hrs 36 min	94
September	164,799	57	356 hrs 55 min	50
October	276,553	57	636 hrs 23 min	86
November	258,024	55	643 hrs 46 min	89
December	296,097	55	723 hrs 26 min	97
Total	3,293,511	55	7,416 hrs 4 min	84

Table 5.2 Landfill Gas Destruction Summary (2020)

Month	Volume LFG Flared (m³)	Mass Methane Destroyed (tonnes GWP 25)	Mass Carbon Dioxide Equivalent Destroyed (tonnes)
January	317,093	109	2,732
February	232,864	84	2,093
March	307,972	109	2,722
April	217,408	78	1,958
May	174,244	62	1,554
June	330,800	118	2,959
July	327,717	116	2,894
August	343,980	115	2,876
September	161,503	60	1,496
October	271,022	101	2,521
November	252,864	91	2,281
December	290,175	105	2,617
Total	3,227,641	1,148	28,703

Table 6.1 Summary of Scheduled Maintenance Shutdowns

Date	Duration (hours)	Reason	Comments
February 3, 2020	48 min	Monthly Maintenance	
February 19, 2020	46 min	CH4 and O2 Analyzer Calibration	Lakewood Electric
February 26 and 27, 2020	9 hrs 35 min	Finish Fire Alarm Testing, Compressor Maintenance	All Points Fire Protection, NW Compressed Air
March 18, 2020	40 min	Monthly Maintenance, CH4 and O2 analyzer calibration	Lakewood Electric
May 20 to May 31, 2020	272 hrs 20 min	Decommissioning of LFG System for the LFG Header Extension Construction Project, Monthly Maintenance	
June 1, 2020	15 hrs 12 min	Recommissioning of the LFG System Upon Completion of the LFG Header Extension Project	
June 3, 2020	18 min	Training staff on Basic Plant Restart Procedure	
June 9, 2020	12 min	Monthly Maintenance, Quarterly and Semi Annual Maintenance	
July 9 and July 10, 2020	23 hrs 6 min	Bater Electric Servicing of Blowers	Bater Electric
July 16, 2020	53 min	CH4 and O2 Analyzer Calibration	Lakewood Electric
August 17, 2020	1 hr	Monthly Maintenance	
September 3 to September 15, 2020	290 hrs 44 min	Decommissioning of LFG System in Support of the Landfill Remediation Project, Monthly Maintenance	
September 21, 2020	5 min	Switched to Blower 2	
September 22 to September 24, 2020	45 hrs, 59 min	System switched off to Complete Tasks Related to the Landfill Remediation Project	
October 8, 2020	1 hr 5 min	CH4 and O2 Analyzer Calibration	Lakewood Electric
October 22, 2020	2 hrs 11 min	Monthly Maintenance, Quarterly Maintenance, Start Semi Annual, Annual Maintenance, Change Out Flow Meter, Environmental Sensor Calibration	Lakewood Electric
November 5, 2020	2 hrs 38 min	Raising of EW24T	
November 18 and 19, 2020	16 hrs 53 min	Finish Annual, Semi Annual Maintenance	
December 1, 2020	23 min	Annual Fire Alarm Inspection Started	All Points Fire Protection
December 3, 2020	42 min	Monthly Maintenance	
December 21 and 22, 2020	19 hrs 29 min	Re calibration of O2 Analyzer, Start Environmental Sensor Replacement	Lakewood Electric

Table 6.2 Summary of Unscheduled Shutdowns

Date	Duration (hours)	Reason	Comments
February 10 to 13, 2020	69 hrs 25 min	Wildlife Flew Into Power Line	
February 15 to 18, 2020	72 hrs 55 min	Suspected Power Surge	BC Hydro
March 17, 2020	2 hrs 20 min	Power Surge	BC Hydro
April 12 to 16, 2020	87 hrs 49 min	Wildlife Flew Into Power Line	
April 23 to 28, 2020	121 hrs 48 min	Power Surge	BC Hydro
May 14 to 16, 2020	45 hrs 08 min	Wildlife Flew Into Power Line	
July 11, 2020	5 hrs 13 min	Wildlife Flew Into Power Line	
July 22 and 23, 2020	23 hrs 34 min	Power Surge/Outage	BC Hydro
August 15 to 17, 2020	47 hrs 24 min	Power Surge/Outage	BC Hydro
September 29 and 30, 2020	26 hrs 17 min	Wildlife Flew Into Power Line	
October 5 to 7, 2020	40 hrs 31 min	Wildlife Flew Into Power Line	
October 16 and 17, 2020	31 hrs 12 min	Power Surge	BC Hydro
October 26 to 28, 2020	32 hrs 38 min	Power Surge/Outage	BC Hydro
November 2, 2020	5 hrs 12 min	O2 Analyzer	Lakewood Electric
November 4, 2020	4 hrs 18 min	Wildlife Flew Into Power Line	
November 10 to 12, 2020	47 hrs 25 min	Power Surge/Outage	BC Hydro

APPENDIX A

METHANE AND OXYGEN ANALYZER CALIBRATION CERTIFICATES

Certificate of Calibration

Report number LE-RDFFG-20200219-0002



2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

Manufacturer	Model	Gauge Number	Serial Number	Calibration Date	Expiration Date
Galvanic IR	IR Gas Analyser	CH4	0249IR011	2019-02-19	2019-08-19

Model Uncertainty
"±/- 1% Span"

All instrument calibrations are verified for accuracy before they are shipped. The recommended calibration interval for this instrument is 6 months from the date of verification. Your particular quality assurance requirements may supersede this recommendation.

As Received Condition: In Tolerance

As Left Condition: In Tolerance

All calibrations are performed in a controlled or operational environment by qualified personnel using instrumentation and methods which guarantee that specifications claimed are reliable. Calibrations conform to ANSI/NC SL Z540-1-1994, MIL-STD 45662A, 10CFR21 and 10CFR50 when specified by customer documentation.

Definitions:	Temperature	Measured temperature of test during data collection.
	Reference Reading	True value according to our reference standards.
	Gauge Reading	Displayed reading from test unit.
	Condition	Pass or Fail.
	Difference	Indicated reading minus reference reading.
	Relative Difference	(Difference / reference reading) x 100
	Allowable Tolerance	± according to manufacturer's specifications.
	Water column	Referenced at 20° C and 1 atmosphere.
	Test Accuracy Ratio	At least 4:1 unless otherwise stated.

Media used during this calibration were:

Manufacture	Praxair
Zero	100% Nitrogen
Linearity	50% Methane
Balance	Nitrogen

Reference Standards used in this calibration are traceable to the National Institute of Standards and Technology of the United States, through the following report numbers:

Manufacturer	Model	Serial Number	Report Number	Due Date	Reference Uncertainty
Crystal Engineering	IS33-36/3000PSI	2262-432200	48487	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)
Crystal Engineering	IS33-36/300PSI	2535-841542	48487-A	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)

This certificate shall not be reproduced except in full, without written approval.

Laboratory Representative
Darren Heggelund

Quality Representative
Darren Heggelund



mA Test Results for CH4

LE-RDFFG-20200219-0002

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

As Received Test Results

4-20 mA


Reference CH ₄ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.995	0.016	0.005	0.03%	Pass
50.00	12.000	12.015	0.016	0.015	0.09%	Pass

As Left Test Results

4-20 mA

Reference CH ₄ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.995	0.016	0.005	0.03%	Pass
50.00	12.000	12.015	0.016	0.015	0.09%	Pass

mA Error Graph for Gauge # CH4, S/N 0249IR011, Report LE-RDFFG-20200219-0002



— As Received Test Results
— As Left Test Results



Certificate of Calibration

Report number LE-RDFFG-20200318-0002



2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

Manufacturer	Model	Gauge Number	Serial Number	Calibration Date	Expiration Date
Galvanic IR	IR Gas Analyser	CH4	0249IR011	2020-03-18	2020-09-18

Model Uncertainty
"±/- 1% Span"

All instrument calibrations are verified for accuracy before they are shipped. The recommended calibration interval for this instrument is 6 months from the date of verification. Your particular quality assurance requirements may supersede this recommendation.

As Received Condition: In Tolerance

As Left Condition: In Tolerance

All calibrations are performed in a controlled or operational environment by qualified personnel using instrumentation and methods which guarantee that specifications claimed are reliable. Calibrations conform to ANSI/NCSL Z540-1-1994, MIL-STD 45662A, 10CFR21 and 10CFR50 when specified by customer documentation.

Definitions:	Temperature	Measured temperature of test during data collection.
	Reference Reading	True value according to our reference standards.
	Gauge Reading	Displayed reading from test unit.
	Condition	Pass or Fail.
	Difference	Indicated reading minus reference reading.
	Relative Difference	(Difference / reference reading) x 100
	Allowable Tolerance	± according to manufacturer's specifications.
	Water column	Referenced at 20° C and 1 atmosphere.
	Test Accuracy Ratio	At least 4:1 unless otherwise stated.

Media used during this calibration were:

Manufacture	Praxair
Zero	100% Nitrogen
Linearity	50% Methane
Balance	Nitrogen

Reference Standards used in this calibration are traceable to the National Institute of Standards and Technology of the United States, through the following report numbers:

Manufacturer	Model	Serial Number	Report Number	Due Date	Reference Uncertainty
Crystal Engineering	IS33-36/3000PSI	2262-432200	48487	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)
Crystal Engineering	IS33-36/300PSI	2535-841542	48487-A	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)

This certificate shall not be reproduced except in full, without written approval.

Laboratory Representative
Darren Heggelund

Quality Representative
Darren Heggelund



mA Test Results for CH4

LE-RDFFG-20200318-0002

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

As Received Test Results

4-20 mA


Reference CH ₄ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.996	0.016	0.004	0.03%	Pass
50.00	12.000	11.967	0.016	0.033	0.21%	Pass

As Left Test Results

4-20 mA

Reference CH ₄ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.996	0.016	0.004	0.03%	Pass
50.00	12.000	11.967	0.016	0.033	0.21%	Pass

mA Error Graph for Gauge # CH4, S/N 0249IR011, Report LE-RDFFG-20200318-0002



As Received Test Results
As Left Test Results



Certificate of Calibration

Report number LE-RDFFG-20200219-0001

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571



Manufacturer	Model	Gauge Number	Serial Number	Calibration Date	Expiration Date
GE Sensing	XMO2-2H-41	O2	1906	2020-02-19	2020-08-19

Model Uncertainty
"±/- 1% Span"

All instrument calibrations are verified for accuracy before they are shipped. The recommended calibration interval for this instrument is 6 months from the date of verification. Your particular quality assurance requirements may supersede this recommendation.

As Received Condition: In tolerance

As Left Condition: In tolerance

All calibrations are performed in a controlled or operational environment by qualified personnel using instrumentation and methods which guarantee that specifications claimed are reliable. Calibrations conform to ANSI/NCSL Z540-1-1994, MIL-STD 45662A, 10CFR21 and 10CFR50 when specified by customer documentation.

Definitions:	Temperature	Measured temperature of test during data collection.
	Reference Reading	True value according to our reference standards.
	Gauge Reading	Displayed reading from test unit.
	Condition	Pass or Fail.
	Difference	Indicated reading minus reference reading.
	Relative Difference	(Difference / reference reading) x 100
	Allowable Tolerance	± according to manufacturer's specifications.
	Water column	Referenced at 20° C and 1 atmosphere.
	Test Accuracy Ratio	At least 4:1 unless otherwise stated.

Media used during this calibration were:

Manufacture	Praxair
Zero	100% Nitrogen
Linearity	20.9% Oxygen
Balance	Nitrogen

Reference Standards used in this calibration are traceable to the National Institute of Standards and Technology of the United States, through the following report numbers:

Manufacturer	Model	Serial Number	Report Number	Due Date	Reference Uncertainty
Crystal Engineering	IS33-36/3000PSI	2262-432200	48487	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)
Crystal Engineering	IS33-36/300PSI	2535-841542	48487-A	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)

This certificate shall not be reproduced except in full, without written approval.

Laboratory Representative
Darren Heggelund

Quality Representative
Darren Heggelund



mA Test Results for O₂

LE-RDFFG-20200219-0001

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

As Received Test Results

4-20 mA


Reference O ₂ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.965	0.640	0.035	0.22%	PASS
20.90	17.376	17.456	0.640	0.080	0.50%	PASS

As Left Test Results

4-20 mA

Reference O ₂ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	4.011	0.640	0.011	0.07%	Pass
20.90	17.376	17.456	0.640	0.080	0.50%	Pass

mA Error Graph for Gauge # O2, S/N 1906, Report LE-RDFFG-20200219-0001



As Received Test Results
As Left Test Results



Certificate of Calibration

Report number LE-RDFFG-20200318-0001

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571



Manufacturer	Model	Gauge Number	Serial Number	Calibration Date	Expiration Date
GE Sensing	XMO2-2H-41	O2	1906	2020-03-18	2020-09-18

Model Uncertainty
"± 1% Span"

All instrument calibrations are verified for accuracy before they are shipped. The recommended calibration interval for this instrument is 6 months from the date of verification. Your particular quality assurance requirements may supersede this recommendation.

As Received Condition: In tolerance

As Left Condition: In tolerance

All calibrations are performed in a controlled or operational environment by qualified personnel using instrumentation and methods which guarantee that specifications claimed are reliable. Calibrations conform to ANSI/NCSL Z540-1-1994, MIL-STD 45662A, 10CFR21 and 10CFR50 when specified by customer documentation.

Definitions:	Temperature	Measured temperature of test during data collection.
	Reference Reading	True value according to our reference standards.
	Gauge Reading	Displayed reading from test unit.
	Condition	Pass or Fail.
	Difference	Indicated reading minus reference reading.
	Relative Difference	(Difference / reference reading) x 100
	Allowable Tolerance	± according to manufacturer's specifications.
	Water column	Referenced at 20° C and 1 atmosphere.
	Test Accuracy Ratio	At least 4:1 unless otherwise stated.

Media used during this calibration were:

Manufacture	Praxair
Zero	100% Nitrogen
Linearity	20.9% Oxygen
Balance	Nitrogen

Reference Standards used in this calibration are traceable to the National Institute of Standards and Technology of the United States, through the following report numbers:

Manufacturer	Model	Serial Number	Report Number	Due Date	Reference Uncertainty
Crystal Engineering	IS33-36/3000PSI	2262-432200	48487	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)
Crystal Engineering	IS33-36/300PSI	2535-841542	48487-A	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)

This certificate shall not be reproduced except in full, without written approval.

Laboratory Representative
Darren Heggelund

Quality Representative
Darren Heggelund



mA Test Results for O₂

LE-RDFFG-20200318-0001

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

As Received Test Results

4-20 mA


Reference O ₂ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.988	0.640	0.012	0.08%	PASS
20.90	17.376	17.366	0.640	0.010	0.06%	PASS

As Left Test Results

4-20 mA

Reference O ₂ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.988	0.640	0.012	0.08%	Pass
20.90	17.376	17.366	0.640	0.010	0.06%	Pass

mA Error Graph for Gauge # O2, S/N 1906, Report LE-RDFFG-20200318-0001



As Received Test Results
As Left Test Results



Certificate of Calibration

Report number LE-RDFFG-20200318-0002



2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

Manufacturer	Model	Gauge Number	Serial Number	Calibration Date	Expiration Date
Galvanic IR	IR Gas Analyser	CH4	0249IR011	2020-07-16	2021-01-16

Model Uncertainty
"±/- 1% Span"

All instrument calibrations are verified for accuracy before they are shipped. The recommended calibration interval for this instrument is 6 months from the date of verification. Your particular quality assurance requirements may supersede this recommendation.

As Received Condition: In Tolerance

As Left Condition: In Tolerance

All calibrations are performed in a controlled or operational environment by qualified personnel using instrumentation and methods which guarantee that specifications claimed are reliable. Calibrations conform to ANSI/NCSL Z540-1-1994, MIL-STD 45662A, 10CFR21 and 10CFR50 when specified by customer documentation.

Definitions:	Temperature	Measured temperature of test during data collection.
	Reference Reading	True value according to our reference standards.
	Gauge Reading	Displayed reading from test unit.
	Condition	Pass or Fail.
	Difference	Indicated reading minus reference reading.
	Relative Difference	(Difference / reference reading) x 100
	Allowable Tolerance	± according to manufacturer's specifications.
	Water column	Referenced at 20° C and 1 atmosphere.
	Test Accuracy Ratio	At least 4:1 unless otherwise stated.

Media used during this calibration were:

Manufacture	Praxair
Zero	100% Nitrogen
Linearity	50% Methane
Balance	Nitrogen

Reference Standards used in this calibration are traceable to the National Institute of Standards and Technology of the United States, through the following report numbers:

Manufacturer	Model	Serial Number	Report Number	Due Date	Reference Uncertainty
Crystal Engineering	IS33-36/3000PSI	2262-432200	48487	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)
Crystal Engineering	IS33-36/300PSI	2535-841542	48487-A	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)

This certificate shall not be reproduced except in full, without written approval.

Laboratory Representative
Darren Heggelund

Quality Representative
Darren Heggelund



mA Test Results for CH4

RDFFG-20200716-CH4

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

As Received Test Results

4-20 mA


Reference CH ₄ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.994	0.016	0.006	0.04%	Pass
50.00	12.000	11.826	0.016	0.174	1.09%	Fail

As Left Test Results

4-20 mA

Reference CH ₄ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.999	0.016	0.001	0.01%	Pass
50.00	12.000	11.991	0.016	0.009	0.06%	Pass

mA Error Graph for Gauge CH4, SN 0249IR011, Report RDFFG-20200716-CH4



As Received Test Results
As Left Test Results



Certificate of Calibration

Report number LE-RDFFG-20200318-0001

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571



Manufacturer	Model	Gauge Number	Serial Number	Calibration Date	Expiration Date
GE Sensing	XMO2-2H-41	O2	1906	2020-07-16	2021-01-16

Model Uncertainty
"± 1% Span"

All instrument calibrations are verified for accuracy before they are shipped. The recommended calibration interval for this instrument is 6 months from the date of verification. Your particular quality assurance requirements may supersede this recommendation.

As Received Condition: In tolerance

As Left Condition: Out of Tolerance

All calibrations are performed in a controlled or operational environment by qualified personnel using instrumentation and methods which guarantee that specifications claimed are reliable. Calibrations conform to ANSI/NCCL Z540-1-1994, MIL-STD 45662A, 10CFR21 and 10CFR50 when specified by customer documentation.

Definitions:	Temperature	Measured temperature of test during data collection.
	Reference Reading	True value according to our reference standards.
	Gauge Reading	Displayed reading from test unit.
	Condition	Pass or Fail.
	Difference	Indicated reading minus reference reading.
	Relative Difference	(Difference / reference reading) x 100
	Allowable Tolerance	± according to manufacturer's specifications.
	Water column	Referenced at 20° C and 1 atmosphere.
	Test Accuracy Ratio	At least 4:1 unless otherwise stated.

Media used during this calibration were:

Manufacture	Praxair
Zero	100% Nitrogen
Linearity	20.9% Oxygen
Balance	Nitrogen

Reference Standards used in this calibration are traceable to the National Institute of Standards and Technology of the United States, through the following report numbers:

Manufacturer	Model	Serial Number	Report Number	Due Date	Reference Uncertainty
Crystal Engineering	IS33-36/3000PSI	2262-432200	48487	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)
Crystal Engineering	IS33-36/300PSI	2535-841542	48487-A	Dec. 11, 2020	+/- (0.05% of Reading + 0.005% of F.S.), vacuum: +/- (0.25% of Reading + 0.004 PSI)

This certificate shall not be reproduced except in full, without written approval.

Laboratory Representative
Darren Heggelund

Quality Representative
Darren Heggelund



mA Test Results for O2

RDFFG-20200716-O2

2007 Ogilvie St., Prince George BC V2N 1X2 | Phone: 250-563-4100 Fax: 250-563-1571

As Received Test Results

4-20 mA


Reference O ₂ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	4.185	0.640	0.185	1.16%	FAIL
20.90	17.376	17.366	0.640	0.010	0.06%	PASS

As Left Test Results

4-20 mA

Reference O ₂ (%/vol)	Calculated mA	Reference mA	Tolerance mA	Difference mA	Difference (% of Span)	Condition
0.00	4.000	3.987	0.640	0.013	0.08%	Pass
20.90	17.376	17.366	0.640	0.010	0.06%	Pass

mA Error Graph for Gauge O2, SN 1906, Report RDFFG-20200716-O2



— As Received Test Results
— As Left Test Results



INFRA-RED BASED MEASUREMENT TECHNOLOGY

GALVANIC APPLIED SCIENCES NOW OFFERS CO₂, CH₄ AND CO MEASUREMENT BASED ON INFRA-RED TECHNOLOGY.



Class 1 Div I IR

The sensor is based on true dual wavelength infra-red technology with no moving parts. The result is a low drift, high accuracy with a fast response time and low power consumption.



General Purpose

PRINCIPLE OF OPERATION

Many gases including CO₂, CO and CH₄ absorb energy in the infra-red spectrum. This absorption is selective, occurring at specific frequencies of bond vibrations within the molecule. Measuring at a characteristic absorption wavelength enables the detection of the gas, and the strength of the absorption gives a measure of the gas concentration.


The infra-red source emits a broad band of radiation that is guided through the gas in the sample cell. The radiation then passes through either of the two filters. The two filters are selected so that one is coincident with the gas to be measured and the other is independent of any gas likely to be present in the cell. In the absence of gas, the energy reaching the two detectors is equal. As the sample gas concentration increases, the amount of infra-red radiation reaching the two detectors differs due to the absorption at one filter wavelength only.



Electronic signal ratioing using a dedicated on-board microprocessor gives the gas concentration, automatically compensated for source intensity changes or contamination of the optical components

APPLICATIONS

- CO₂ in natural gas
- CH₄ in land fill gas
- Food storage and packaging
- Environmental



WWW.GALVANIC.COM



SPECIFICATIONS

Measuring Range:

CO ₂	CH ₄	CO
0-2000 ppm	0-10%	0-10%
0-3000 ppm	0-30%	
0-5000 ppm	0-100%	
0-1%		
0-3%		
0-10%		
0-100%		

Accuracy:

+/- 2% of range

Stability:

+/- 2% of range per year

Repeatability:

At zero +/- 0.3%
At span +/- 1.5%

Response Time:

T90=10 seconds

Operating Temperature:

0-45 C
(temperature compensated)

Output:


4-20 ma
Lcd digital display
3 relays

Power:

24 vdc 6 watts (standard)
120/240 vac (optional)




Electrical:

General area classification
(standard)
Class 1 Div I Grps C, D
(optional)



Dual Wavelength
IR Sensor

DIMENSIONS



ISO 9002:1994
Registered QMS

Galvanic Applied Sciences Inc.
Bay 6, 6325-12 Street S.E.
Calgary, Alberta T2H 2K1
Telephone: 403-252-8470
Fax: 403-255-6287

Calgary, Alberta
1200, 715-5th Avenue S.W.
Calgary, Alberta T2P 2X6
Telephone: 403-252-8470
Fax: 403-571-2684

Edmonton, Alberta
Telephone: 780-430-0142
Fax: 780-430-0172

Houston, Texas
Telephone: 281-477-9002
Fax: 281-477-9004

Toll Free 1-866-252-8470 (US & Canada only)

WWW.GALVANIC.COM • E-MAIL: INFO@GALVANIC.COM

GE Sensing

Applications

An oxygen transmitter for use in:

- Inerting/blanketing liquid storage tanks
- Reactor feed gases
- Centrifuge gases
- Catalyst regeneration
- Solvent recovery
- Landfill gas
- Sewage wastewater digester gas
- Oxygen purity

Features

- Measurement ranges from 0.01 percent to 100 percent O₂ in gases
- Explosion-proof and flameproof enclosure with weatherproof protection allows sensor to be remotely mounted at the measurement point
- Push-button, single or dual gas calibration
- Compact, rugged sensor design with no moving parts provides long term reliability and trouble-free operation
- Dual-bridge measurement circuit compensates for variations in background gas composition
- Unique dual-chamber, temperature-controlled cell design provides resistance to contamination and flow fluctuation
- Computer-enhanced accuracy of 1 percent of span and linearity of better than 0.5 percent of span

XMO2 Panametrics Smart Oxygen Analyzer

XMO2 is a Panametrics product. Panametrics has joined other GE high-technology sensing businesses under a new name—GE Industrial, Sensing.



GE Sensing



XMO2 rack-mount configuration

Smart Oxygen Transmitter

The XMO2 thermoparamagnetic oxygen transmitter is the most stable oxygen analyzer available on the market today. It represents the state of the art in oxygen measurement. With the XMO2 transmitter, reliable process oxygen measurement can be as easy as temperature or pressure measurement.

Top Performance and Ease of Use

The XMO2 combines computer-enhanced, automatic oxygen signal compensation, fast-response software, real-time error detection and automated calibration with a proven thermoparamagnetic oxygen sensor to achieve unequalled performance and ease of use.

The compact, weatherproof, explosion-proof and flameproof XMO2 is specifically designed to be field installed at the process measurement point, thus minimizing sample-conditioning requirements while ensuring the best sample and the fastest possible response. With no moving parts, it is insensitive to mounting position or vibration, and it has excellent long-term reliability. The XMO2's dual-chamber oxygen cell design makes it resistant to contamination and flow variation.

Automatic Background Gas Compensation

An onboard microprocessor gives the XMO2 the computing power to provide advanced online signal conditioning and digital communications via an RS232 interface and menu-driven software. Integrated signal-processing algorithms provide improved linearity and accuracy, and automatic compensation for background gas variations and/or atmospheric pressure effects. A fast-response software routine provides typical response times of less than 15 seconds. When recalibration does become necessary, it can be accomplished quickly and easily through software, with no potentiometers to adjust.


Choice of Enclosures and Ranges

The XMO2 requires 24 VDC power and provides a 4 to 20 mA output signal with fully programmable zero and span settings. The output is proportional to oxygen concentration and internally compensated for background gas and/or atmospheric pressure variations. The weatherproof, explosion-proof and rack-mount XMO2s are available in a wide variety of measurement ranges.

XMO2 Accessories

GE provides a complete line of accessories for use with the XMO2. This includes sample systems custom-designed for specific applications, a 24 VDC power supply and a four-wire color-coded cable in lengths up to 4000 ft (1200 m). The XMO2 can also be interfaced with other GE displays and analyzers, such as the TMO2D, XDP and Moisture Series analyzers. The TMO2D and XDP displays provide microprocessor-based oxygen signal compensation for maximum accuracy, software-enhanced response and automatic calibration of the XMO2 transmitter.

Dual-Chamber Design



Flow schematic of the XMO2 thermoparamagnetic oxygen measuring cell. Oxygen's paramagnetic property causes an oxygen-containing gas sample to move within the magnetic field. The gas movement creates a "magnetic wind" that is sensed by the thermistor pairs. Oxygen concentration and background gas compensation are determined by the transmitter's microprocessor.

XMO2 Specifications

Performance

Accuracy

- ±1% of span
- ±2% of span for 0 to 1% range
- ±0.2% O₂ for 90 to 100% and 80 to 100% ranges

Linearity

±0.5% of span

Repeatability

±0.2% of span

Measurement Resolution

0.01 mA

Zero Stability

±1% of span per month (±2% for 0 to 1% range)

Span Stability

±0.4% of span per month (±0.8% for 0 to 1% range)

Measurement Ranges (Typical)

- 0% to 1%
- 0% to 2%
- 0% to 5%
- 0% to 10%
- 0% to 21%
- 0% to 25%
- 0% to 50%*
- 0% to 100%*
- 90% to 100%*
- 80% to 100%*


* Pressure compensation required

Transmitter Temperature

- Standard: Controlled to 113°F (45°C)
- Optional: Controlled to 140°F (60°C)



The XMO2 output may be used as an input to GE Moisture Series analyzers for simultaneous measurement and display of both moisture and oxygen content.



XMO2 oxygen transmitter dimensions

Pressure Effect

- ±0.2% of reading per mm Hg (without pressure compensation)
- Option available for pressure compensation

Required Sample Flow Rate

0.1–2.0 SCFH (50–1,000 cc/min),
1.0 SCFH (500 cc/min) nominal

Sample Flow Rate Effect

Less than 1% of span for flow range of 0.1 to 2.0 SCFH (50 to 1000 cc/min) for weatherproof XMO2 with background gas compensation

Response Time, 90% Step Change

- Fast 15 seconds
- EN50104 45 seconds
- Standard 70 seconds

Warm-Up Time

30 minutes

Functional

Analog Output

4 to 20 mA, isolated, 800 Ω maximum, field programmable

GE Sensing

Digital Output
RS232, three-wire

Power
24 VDC ±4 VDC, 1.2 A maximum

Cable

- Standard: 10 ft (3 m), four-wire
- Lengths up to 4000 ft (1200 m) available for current output

Ambient Temperature Range
(Sample Conditions):

- -4°F to 104°F (-20°C to 40°C), standard cell operating temperature of 113°F (45°C)
- 23°F to 131°F (-5°C to 55°C), optional cell operating temperature of 140°F (60°C)

Maximum Pressure
20 psig (2 bar)

Physical

Wetted Sensor Materials

- Standard: 316 stainless steel, glass and Viton® O-rings
- Optional: Hastelloy® C276 and Chemraz® O-rings

Dimensions

- Weatherproof unit (h x dia): 9.53 in x 5.71 in (242 mm x 145 mm)
- Explosion-proof/weatherproof unit (h x dia): 10.47 in x 5.71 in (266 mm x 145 mm)

Weight
9.5 lb (4.3 kg)

Environmental, Transmitter

- Weatherproof: Type 4X/IP66
- Explosion-proof: Class I, Division 1, Groups A,B,C&D, FM/CSA
- Flameproof:
⊕ II 2 GD EEx d IIC T6
ISseP02ATEX022;
Ex d II C T6 IP66 Class I, Zone 1
SAA AUS Ex 3139X

ATEX compliance with EN50104 requires response-time calibration to EN50104 and constant control of sample-system pressure or pressure compensation of XMO2.

Environmental, Rack Mount
Rack-mount configuration is suitable for ordinary locations only. Not suitable for use in hazardous (classified) locations.

European Compliance
Complies with EMC Directive 89/336/EEC and PED 97/23/EC for DN<25 (CE approval pending for rack mount)

Lloyd's Registry Approval
Please refer to the XMO2-LR data sheet for details.

Order Information

Record selected option in blank indicated at bottom of form.

XMO2 Thermomagnetic Oxygen Transmitter

Package

- 1 Weatherproof enclosure
- 2 Explosion-proof/weatherproof enclosure
- 5 Rack-mount configuration
- X Without enclosure (spare)

Cell Magnetization

- H High (suitable for 0 to 1%, 0 to 2%, 0 to 5%, 0 to 10%, 0 to 21%, 0 to 25%, 0 to 50%, 90 to 100%, 80 to 100% and 0 to 100% ranges)

Compensation

- 3 Background gas only (standard)
- 4 Atmospheric pressure and background gas (optional)

Material

- 1 316 stainless steel
- 2 Hastelloy C276

XMO2 - _ _ _ _ Use this number to order product

XMO2 Calibration Specifications

Range of Oxygen Output

- | | | |
|------------|--------------|---------------|
| 1 0 to 1% | 5 0 to 21% | A 90 to 100%* |
| 2 0 to 2% | 6 0 to 25% | B 80 to 100%* |
| 3 0 to 5% | 7 0 to 50%* | S Special |
| 4 0 to 10% | 8 0 to 100%* | |

Compensation Signal

- 1 Background gas, standard N₂/CO₂
- 2 Atmospheric pressure, standard range (700 to 800 mm Hg)

Response

- 1 Standard response
- 2 Response to meet EN50104
- 3 Fast response

XCAL - _ _ _ _ Use this number to order product

*Requires pressure compensation option.



©2005 GE. All rights reserved.
920-032B

All specifications are subject to change for product improvement without notice. GE® is a registered trademark of General Electric Co. Other company or product names mentioned in this document may be trademarks or registered trademarks of their respective companies, which are not affiliated with GE.



www.gesensing.com

APPENDIX B

FLOW METER CALIBRATION CERTIFICATE

SERIAL NUMBER 2008045

02-27-2020

COMPONENT VALUES DETERMINED AT CALIBRATION

PCB 100287

PB 100301

R3 100
R4 455
R11 5900
R18 800
R15

R9 49900
R10 7320
R11 49900
R13 7320
DP 56200

FLOW SENSOR - 50 OHMS NI @75 °F
TEMP SENSOR - 100 OHMS NI @75 °F

CALIBRATION POTENTIOMETERS

TA 765

TC 592


OPERATING RANGE
minimum maximum

FLOW	0	750	SCFM LFG
TEMPERATURE	33	200	°F
PRESSURE	0	15	PSIG

COMMENTS: 6.357" LINE ID
ONE PULSE = 500 SCF

METER OUTPUT SIGNAL CALIBRATION

mAmps	SCFM LFG	EXC Volts
4.00	0.00	7.895
5.60	75.00	8.819
7.20	150.00	9.360
8.80	225.00	9.740
10.40	300.00	10.031
12.00	375.00	10.263
13.60	450.00	10.456
15.20	525.00	10.619
16.80	600.00	10.761
18.40	675.00	10.887
20.00	750.00	10.999




FIGURE

TRANSDUCER VOLTAGE

PAGE

SCFM LFG


S/N 2008045.0
02-27-2020



FIGURE

OUTPUT CURRENT (mA)

PAGE



S/N 2008045 Temperature Degrees F

Thermal Instrument Company

217 Sterner Mill Road, Trevoise, PA 19053
 Telephone No. (215) 355-8400 Fax No. (215) 355-1789
 Email: Office@Thermalinstrument.com Web: www.Thermalinstrument.com

Calibration Test Data Sheet

SN: 2008045	Master Meter: Venturi S/N 75279	Fluid: Landfill Gas
Flow Range: 0 - 750 SCFM	Temperature Range: 33 - 200 Deg F	Pressure Range: 0 - 15 PSIG
Accuracy: +/- 1% FS Failed	Calibrator: BB	Line Size: 6.357" Line ID
Date Calibrated: 2/26/2020	Customer: Regional District of Fraser- Fort George	Approved By: <u>Susan Roxberry</u>

As Found

Master Meter Indication	% Flow Range	Test Meter Exc. (DC V)	Test Meter Output #1 (DC mA)	% MA Range	% Error	Test Meter Output #2 (If Avail)	% Display Range	% Display Error
0	0	7.895	4.148	0.9	0.9	0	0	0
90.8	12.1	8.953	6.06	12.9	0.8	0	0	0
168.4	22.5	9.464	7.712	23.2	0.7	0	0	0
221.3	29.5	9.724	8.846	30.3	0.8	0	0	0
327.5	43.7	10.122	11.116	44.5	0.8	0	0	0
368.7	49.2	10.245	12	50	0.8	0	0	0
466.2	62.2	10.494	14.098	63.1	1	0	0	0
549.1	73.2	10.667	15.875	74.2	1	0	0	0
615.9	82.1	10.789	17.279	83	0.9	0	0	0
662.5	88.3	10.866	18.303	89.4	1.1	0	0	0

Thermal Instrument Company

217 Sterner Mill Road, Trevoise, PA 19053
 Telephone No. (215) 355-8400 Fax No. (215) 355-1789
 Email: Office@Thermalinstrument.com Web: www.Thermalinstrument.com

Calibration Test Data Sheet

SN: 2008045	Master Meter: Venturi S/N 75279	Fluid: Landfill Gas
Flow Range: 0 - 750 SCFM	Temperature Range: 33 - 200 Deg F	Pressure Range: 0 - 15 PSIG
Accuracy: +/- 1% FS	Calibrator: BB	Line Size: 6.357" Line ID
Date Calibrated: 2/27/2020	Customer: Regional District of Fraser-Fort George	Approved By: <u>Susan Rothberry</u>

As Left

Master Meter Indication	% Flow Range	Test Meter Exc. (DC V)	Test Meter Output #1 (DC mA)	% MA Range	% Error	Test Meter Output #2 (If Avail)	% Display Range	% Display Error
0	0	7.895	4.069	0.4	0.4	0	0	0
88.1	11.7	8.931	5.873	11.7	0	0	0	0
132.6	17.7	9.253	6.802	17.5	-0.2	0	0	0
248.3	33.1	9.838	9.24	32.7	-0.4	0	0	0
279.4	37.3	9.958	9.905	36.9	-0.3	0	0	0
381.7	50.9	10.282	12.084	50.5	-0.4	0	0	0
466.9	62.3	10.495	13.915	62	-0.3	0	0	0
501.2	66.8	10.57	14.659	66.6	-0.2	0	0	0
628.4	83.8	10.81	17.363	83.5	-0.3	0	0	0
662.5	88.3	10.866	18.135	88.3	0	0	0	0

CALIBRATION CERTIFICATION

We certify that the calibration accuracies listed below are obtained on equipment, and with methods, that can be traced directly to the US National Institute of Standards and Technology.

FLOWRATE READOUT ACCURACY: \pm 1% Full Scale

PRESSURE TESTED AT:

METER SERIAL NUMBER: **2008045**

MODEL NUMBER: **62-9/926**

The calibration listed above was performed under the following conditions:

0 to 750 SCFM LANDFILL GAS
33 to 200 DEG F
0 to 15 PSIG
6.357" LINE ID

Standard Conditions: 60 Deg. F @ 1 ATM

Signature Brian A. Boyd Date **February 27, 2020**

THERMAL INSTRUMENT COMPANY, INC

217 Sterner Mill Road, Trevoese, PA 19053 Phone: 215-355-8400 Fax: 215-355-1789 Web: www.thermalinstrument.com

Flowmeters by.....

Thermal Instrument Company, Inc

217 Sterner Mill Road, Trevoze, PA 19053
(215) 355-8400 FAX (215) 355-1789
Website: www.thermalinstrument.com
e-mail: office@thermalinstrument.com

MODEL 62-9/62T-9/9500P INTEGRAL THERMAL MASS PROBE FLOWMETER

The Thermal Instrument Model 62-9 Mass Probe Flowmeter With Integral 9500P Microprocessor Electronics brings an element of versatility to Mass Flow Measurement of Liquids and Gases.

A uniquely designed microprocessor controlled electronics provides integral local control and display in an explosion proof housing. Optional Remote Mounting of the electronics up to a distance of 4000 ft. provides additional versatility depending on the application.

Flow signal linearity and temperature effects, determined at calibration, are corrected by the microprocessor using operating parameter data stored in EEPROM memory.

DESIGN FEATURES

- > **Unobstructed Flow Tube with Non-wetted Sensors** mounted on the Inner Surface of the flow tube of the Insertion Probe Design permits measurement of Gases, Liquids and Slurries.
- > **Wide Selection of Probe Materials:** 316 SS is standard. Corrosion resistant materials such as Hastelloy C, Monel, Inconel, Carpenter 20 as well as carbide and fluorocarbon coatings are available.
- > **Hi Temperature designs to 800 F.**
- > **Wide Flow Range on Gases** is greater than 100:1
- > **Vent Flow Measurement** at near atmospheric conditions and low flow.
- > **Rate and Totalizer** Ultra-bright LED display
- > **NIST Traceable Calibration** on user gases and liquids. Calibration in user piping configurations.

SPECIFICATION MODEL 62-9/9500P



- > **Isolated Fluid Temperature Signal**, 4-20MA from an RTD Sensor on the Flow Tube.
- > **Multi sensor Design** for large ducts and stacks
- > **Electronics suitable for Remote Mount** to 4000 ft.
- > **Totalizer Data** is stored in the EEPROM memory on power loss and restored on startup. No batteries required.

ENGINEERING SPECIFICATIONS

Accuracy:

+/- .50% of Full Scale, [or 2.0% of reading whichever is better. Full Scale over the calibrated range.]
[Process fluid temperature span +/- 50 F]

Repeatability: +/- .20 % of Rate

Response Time:

Gases: 1 to 2 sec. Typical
Liquids: Less than 500ms.

Pressure Rating:

0 - 1200 psig (Std)

0 - 10000 psig Dependent on meter size

Mass Flowrates:**Gas:** 40 to 50,000 SFPM**Liquid:** .02 to 20 FPS

Note: maximum determined by pipe capacity.

Flow Range: 100:1 Std. Multiple ranges available.**Calibrated Temperature Range:****Gas:** -40 F to 350 F (-40 to 177 C) STD.

[-20 F to 800 F] (-28 C to 427 C) Optional).

Liquid: -40 F to 350 F (-40 to 177 C) STD.

[-40 F to 500 F](-40 to 260 C) Optional

Calibration Temperature Capability:

-200 TO 800 F (-129 to 427 C)

Wetted Materials: 316SS (STD). Consult the factory for other material requirements.**Mounting:** Flange, Packing Gland, Triclover, Hot Tap**MICROPROCESSOR SPECIFICATIONS:****EEPROM Memory:** 64K**Displays:** 8 digit Flowrate and Totalizer (Optional)**Communicator:****Interface:** RS-232 via Ribbon Connector**Baud Rate:** 9600**Adjustable Variables with RS-232 Communicator:****Zero Cutoff; Zero Offset**

Filter Factor

Flow Full Scale

Flow Rate Decimal Point

Flow Totalizer Decimal Point

Flow Totalizer/Flowrate Ratio

Power: 120/240 vac, 50/60 Hz.

24VDC, 1 Amp (Factory Set)

Electrical Connections: 3/4" FNPT Power & Signal**Enclosure:** Cast Aluminum, Epoxy Painted**Environmental Ratings:**

Housing: Explosion Proof, Class 1, Div. 1,

Groups, B, C D Weatherproof: Nema 4

Ambient Temperature: -25 to 60 C (-13 to 140 F)

Storage Temperature: -40 to 70 C (-40 to 158F)

Mounting:

Integral to the Flow Transducer.

Remote to 4000 ft., Wall or Frame mount.

Signal Cable for Remote Design: 18 gauge, 2 pair with cable shield. Maximum distance, 4000 ft.**MODEL NUMBERING SYSTEM:****Base Model:** = 62-9/9500P**Electronics MOUNTING:****I** = Integral Mount**R** = Remote Mount**Fluid:****G** = Gas**L** = Liquid**Probe Size:** 1/2 to 2" (See Sizing Tables for Gas and Liquid).**Material:****316SS** = STD.**Other** = Consult Factory**Process Connection:****PG** = Packing Gland (Std.)**FLG-150** = 150# RF Flange (Std. Van stone)**TC** = Triclover**FLG-Other** = Consult Factory**Other** = Consult Factory**Power:****24** = 22-26 VDC, 1AMP. 4 WIRE**120** = 120VAC, 50/60 HZ (STD.)**240** = 240VAC, 50/60 HZ**Output Signals: (Select One**)****0/5** = 0-5 VDC****0/10** = 0-10 VDC****4/20** = 4-20MADC**into 750 ohms, Isolated**4/20T** = **Optional** Process Fluid Temp. **4-20MA. Self Powered****Display: (Optional)****D** = Flowrate Display, 5 1/2 Digit**8T** = Totalizer**D/8T** = Flowrate and Total Display, 8 digit**Signal Cable for Remote Design:****SC** = 25 FT Supplied with remote configuration

Thermal Instrument Company, Inc

(215) 355-8400 FAX (215) 355-1789

Website:www.thermalinstrument.com

e-Mail:office@thermalinstrument.com

APPENDIX C

CALCULATION SUMMARY

FOOTHILLS BOULEVARD LANDFILL
GREENHOUSE GAS EMISSIONS REPORT
CALCULATION SUMMARY

Total Average Daily Methane Concentration

$$\bar{c}_{\text{daily}} = (c_x) / (\sum x)$$

Where,

\bar{c}_{daily} - Daily average methane concentration (volumetric basis).

c_x - Recorded methane concentration (volumetric basis).

$\sum x$ - Total number of data points recorded during the day.

Total Average Weekly Methane Concentration

$$\bar{c}_{\text{weekly}} = (\sum \bar{c}_{\text{daily}}) / (\sum x)$$

Where,

\bar{c}_{weekly} - Weekly average methane concentration (volumetric basis).

\bar{c}_{daily} - Daily average methane concentration (volumetric basis).

$\sum x$ - Total number of days data points were recorded during the week.

Methane Destruction

$$M = \rho_m (\sum c_x V_x) (DE)$$

Where,

M - Mass methane destroyed.

ρ_m - Density of methane at 25°C (0.6557 tonnes per cubic metre).

c_x - Methane concentration (volumetric basis).

V_x - Volume of landfill gas destroyed (cubic metres).

DE - Destruction efficiency (98 percent).

Emission Reduction

$$r = (GWP) \rho_m (\sum c_x V_x) (DE)$$

Where,

r - Unit emission reduction.

(GWP) - global warming potential of 21 for methane.

ρ_m - Density of methane at 25°C (0.0006557 tonnes per cubic metre).

c_x - Methane concentration (volumetric basis).

V_x - Volume of landfill gas destroyed (cubic metres).

DE - Destruction efficiency (98 percent).

APPENDIX D

DATA RECORDS